

INDIAN ASSOCIATION FOR THE CULTIVATION OF SCIENCE

2A & 2B , Raja S.C. Mullick Road, Jadavpur, Kolkata-700032

NIT No. : IACS/SCS/AM/2018-19/58

Date: 04/10/2018

Sealed tender in two bids system (technical bid and price bid) is invited from bonafide, resourceful and eligible manufacturer/exclusive distributor/vendors for **Rotary Evaporator with Diaphragm Pump**.

Part-I (Technical Bid) of the tender should contain technical specifications in detail as well as commercial terms and conditions. Part-II (Price Bid) should clearly indicate group-wise price, if needed, as mentioned in the technical bid. The Technical Bid and Price Bid are to be submitted in separately sealed envelopes, distinctly marked accordingly and both to be put inside another envelop, that should be sealed and superscribed with tender notice no. and due date. The bidders may submit bids duly signed in their own letterheads.

Completed tender bids should reach the office of the **Registrar, Indian Association for the Cultivation of Science (IACS), 2A & 2B Raja S. C. Mullick Road, Jadavpur, Kolkata-700032** on or before the scheduled date and time specified below:

Tender Notice No.	NIT No. IACS/SCS/AM/2018-19/58, Date: 04-10-18
Last date and time of submitting tender	13.11.2018 at 12:30 P.M.
Pre-bid meeting(date, time & place) to discuss technical specification	25.10.2018 at 3 PM in the office of Dr. Amit Majumdar, School of Chemical Sciences, IACS
Date and time of opening tender (Technical Bids)	13.11.2018 at 3 P.M.
Place of opening tender	In the office of Dr. Amit Majumdar, School of Chemical Sciences, IACS
Date and time of opening of Price Bid	The Price Bids of the bidders qualifying the technical bid will only be opened, the date of which will be intimated to the short-listed bidders at their email addresses. The rest of the bids will be rejected.
Contact	Email: icam@iacs.res.in Tel.+91 33 2473 4971 [Ext. 1165]

The technical bids will be opened first to judge/evaluate the technical specifications of the said instrument and thereafter the price bids of only technically qualified bidders will be opened.

Technical Bid Evaluation: The Technical Bids will be evaluated in the presence of the representatives of intending bidders who will be able to clarify technical aspects of their bids, if any, required by the Technical Evaluation Team.

Opening of price-bid: The Price Bids of the bidders qualifying the technical bid will only be opened, the date of which will be intimated to the short-listed bidders at their email addresses. The rest of the bids will be rejected.

Please note that IACS will not provide any accommodation or reimburse any expenses to any of the bidders for attending opening of technical bid.

Quotations received incomplete or beyond the stipulated time will be summarily rejected.

Bidders should submit their past experience for supplying and successful installation of similar units to other research Institutes/Universities/other organizations in India and abroad. Please provide documentary proofs of such successful installation and supportive documents that the instruments are running successfully.

1. TECHNICAL BID

The technical bid should contain technical specifications and should be kept in a separate sealed envelope duly superscribed as 'TECHNICAL BID' on the outer cover of the envelop as already detailed above. **It should be clearly mentioned on the envelope as "Technical Specification for Rotary Evaporator with Diaphragm Pump".**

Technical Specification for Rotary Evaporator with Diaphragm Pump:

Type	1. One number of Rotary Vacuum Evaporator consisting of one number of vacuum pump to meet the essential needs in classical laboratory applications. 2. Large vertical condenser. 3. Easy-clip/combi-clip for fixing and removal of non sticking vapor duct and evaporation flask along with other necessary accessories. 4. Power supply 1400 W or less for saving power. 5. All glass components should be made of borosilicate glass.
Controller	6. Digital controller with large digital LCD display (3.5 inch or more) which indicates rotation speed and heating bath temperature.
Vertical condenser	7. Vertical condenser with cooling surface area of about 1,200-2000 cm². 8. Motorized or manual lift up to 150 mm or more with automatic/manual lifting of evaporating flask. 9. Rate of evaporation = 2000-7500 mL/hour (Acetone) or more.
Heating bath	10. Capacity of 4.5 L or more. 11. Heating-bath temp control : Electronic /Digital 12. Controlled temperature range of 20-95°C or more with set and actual digital temperature display. 13. Temperature regulation accuracy should be $\pm 2^{\circ}\text{C}$ or better. 14. Over-temperature protection power cut off when actual temperature exceeds set temperature or by rapid temperature increase. 15. Heating bath should feature non-slip handles to facilitate carrying, refilling and emptying the heating bath and to protect from scalding bath liquid.
Rotation speed	16. 20-270 rpm or more to be controlled electronically with LED display.

Evaporation flask	17. Should supply at least two evaporating flasks with capacity of 500 mL each & at least two receiving flasks with capacity of 1 L. The receiving flasks should preferably have shock proof coating. 18. Flask size 50 to 4000 mL can be used on the same joint adapter without additional connections.
Sealing system	19. High resistance and durable PTFE seal. 20. Preferably Fluorocarbon elastomer material (FKM) O-ring to resist harsh chemicals for longer life.
Clamping sleeve	21. Clamping sleeve for easy fixing and removal and long lasting vapour duct and evaporation flask.
Woulff Bottle with bracket	22. For additional recovery of solvent to protect vacuum pump with volume of 250 ml. <i>Should be quoted as optional item.</i>
Vacuum pump	23. Double stage Diaphragm Vacuum Pump made from chemically resistant materials such as PTFE. 24. Capacity: 5 to 10 mbar. Suction capacity = 1.5 m³/h or more with ultimate vacuum of 7 mbar or more. 25. Integrated/manual vacuum controller
Vacuum controller	26. Manual/digital vacuum controller, pressure range 0-1000 mbar or more with digital display 3.5 inch or more, time function and overpressure function. <i>Should be quoted as optional item.</i>
Accessories	27. At least two extra clamps for receiving flask.
Installation and Demonstration	28. The system should be installed in lab free of cost. Onsite demonstration and training of the system by the application specialist to our lab personnel.
Compliance	29. All components should be from the same manufacturer and all the specifications should be reflected in printed catalogue.
Warranty	30. At least 3 years warranty.
Supporting Brochure	31. As per international standard, a safety certificate should be provided for “continuous operation and safety in laboratory” from Principal. 32. Should have supporting printed brochure of every specification.
Performance & Service back up	33. List of reputed users (of which at least 5 users in the eastern region), local service support with name, address and contact numbers.

Installation: Free installation and user training

Warranty: 3 years warranty.

Please SUBMIT ITEMIZED QUOTES with separate pricing for each accessory.

Installation/Demonstration/Application training at site: Installation & user training at IACS, free of cost by the supplier.

Service facility: In India, preferably Kolkata, supplier should mention their details of service setup and man powers who are responsible for after sales support. Response time should be within 24 hrs.

A compliance table (see below) must be prepared and submitted along with the technical bid.

Sr. No	Tender specification	Your offered instrument specification	Extent of compliance

Tender will not be accepted if the technical compliance chart is NOT provided following the above format.

2. PRICE BID

The financial bid indicating (item-wise) price for the item(s) mentioned in the technical bid should be kept in a separate sealed envelope duly super scribed as 'PRICE BID' on the outer cover of the envelop as already detailed above. Price bids of only technically qualified bidders will be opened and they will be intimated the date and time of the opening of price bid at their e-mail ids. Rest of the bidders will stand rejected.

PRICE: Price to be quoted on CIF Kolkata and also FOB basis.

3. BID SECURITY:

An Account payee Demand Draft of any of the Commercial Banks for Rs. 20,000 (Rupees Twenty Thousand only) drawn in favour of "Indian Association for the Cultivation of Science (State Bank of India, Jadavpur University Branch, A/C No. 11079699211, IFSC: SBIN000093, MICR Code: 700002048)" is to be furnished by the bidders alongwith Technical Bid, except Micro and Small Enterprises (MSEs) as defined in MSE Procurement Policy issued by Department of Micro, Small and Medium Enterprises (MSME) or are registered with the Central Purchase Organisation or the concerned Ministry or Department.

Alternatively, the bidder shall have the option to furnish EMD in the form of Fixed Deposit Receipt, Banker's Cheque or Bank Guarantee from any of the Commercial Banks. The Bid-Security should have at least 90 (ninety) days validity period beyond the final bid validity period.

In case of non-award of the work the Bid Security money would be returned to the unsuccessful Bidders.

4. PERFORMANCE SECURITY:

An Account Payee Demand Draft on any commercial bank of India or Bank Guarantee (BG) of any Commercial Bank preferably located at Kolkata of 10% of the order value in the name of “Indian Association for the Cultivation of Science” is to be furnished by the successful bidder as Performance security. Performance security money should remain valid for a period of 60 days beyond the date of completion of all contractual obligations of the supplier including warranty obligations. Bid security money or EMD will be refunded to successful bidder on receipt of the Performance security money.

5. TERMS OF PAYMENT :

Payment will be made through irrevocable Letter of Credit in two instalments. 90% of the money will be released on submission of shipping of documents. Remaining 10% will be released after successful installation of the instrument.

6. GENERAL INSTRUCTIONS :

1. Validity of tender: Tender submitted should remain valid for at least six months from the date of opening the tender. Validity beyond six months from the date of opening of the tender shall be lapsed by mutual consent.
2. The tender should accompany a compliance chart.
3. Incomplete and conditional tenders as well as tenders received after the due date will be summarily rejected without assigning any reasons thereof.
4. At any time prior to the bid due date, IACS may, for any reason, whether at its own initiative or in response to a clarification requested by a prospective bidder during pre-bid meeting, modify the bidding documents. The amendment(s) will be notified on the institute website. Prospective bidders are advised to occasionally visit the website (www.iacs.res.in/tender) for any amendment.
5. Installation/Demonstration/Application training at site: Installation & user training at IACS, free of cost by the supplier.
6. Service facility: In India, preferably Kolkata, supplier should mention their details of service setup and man powers who are responsible for after sales support. Response time should be within 24 hrs.
7. The model number, make and a printed literature of the product should be submitted positively.
8. Proposed delivery schedule should be mentioned clearly.
9. Manufacturers / exclusive distributors / vendors should have history of supplying this type of instruments to this or other scientific organizations. Availability of a list in this regard would be preferred.
10. Authorized dealership certificate should be provided in case of principal manufacturing company is not quoting directly.
11. Guarantee certificate, users manuals etc. are to be handed over to the user after successful commissioning of the system.
12. In the event of date being declared a closed holiday for purchaser's office, the due date for submission of bids and opening of the technical bids will be the following working day at the appointed time.
13. In case of any dispute, the decision of IACS authority shall be final and binding on the bidders.
14. In case of any litigation the matter will be dealt within the jurisdiction of the High Court of Calcutta only
15. For any clarification regarding technical specifications, information etc., please send your queries to Dr. Amit Majumdar (icam@iacs.res.in).
16. The authority of IACS reserves the right to reject any or all of the tenders received without assigning any reason thereof.

Acting Registrar