

INDIAN ASSOCIATION FOR THE CULTIVATION OF SCIENCE
2A & 2B, Raja S.C. Mullick Road, Jadavpur, Kolkata-700032

Ref. No. IACS/BC/SSJ-P/2017-38/1

Date : 30-08-2017

Corrigendum Notice No. I

NIT No. IACS/BC/SSJ-P/2017-18/38, Date : 10-08-2017

Technical Specification for Stereo Microscope (Corrigendum)

TRINOCULAR STEREOZOOM MICROSCOPE With Transmitted and reflected led illumination, Diascopic base & High end camera with image analysis software

Microscope should have following features

- a) Zooming ratio (8:1 or higher)
- b) Objectives (1x)
- c) Easy to use for dissecting organisms such as zebrafish, mice embryos
- d) Should have impressive price-performance ratio
- e) Should work in between ambient temperature at +5° to +45° Celsius
- f) Camera should provide image quality ≥ 5 mega pixel and able to perform time lapse imaging
- g) Power supply: 220-240V/50Hz
- h) Integrated transmitted LED/OCC illumination or equivalent

ACTING REGISTRAR