

Corrigendum 1

Tender No.: NIT No. : IACS/CSS/BBD/2017-18/05

Sub: Purchase of " Fluorescent Microscope (Inverted)"

Amendments in Tender specification are highlighted

Features	Specification for Fluorescent Microscope (inverted)
Objectives	Fluorescence grade Objectives: 10X/0.25 PH; <u>20X/0.5</u> , <u>40X/0.55 PH</u> and <u>100x/1.32 PH</u> Oil.
Condenser	Condenser with min. Working Distance of 40 mm or more and should have <u>5 or more positions</u> .
Compatibility	High end Microscope, Camera and software from the same manufacturer for better compatibility. <u>UPS could be from a different vendor</u> .

Change in Tender Submitting: 11/05/2017 upto 1.00PM and tender opening date: 11/05/2017 at 3.00PM

Other terms and conditions as mentioned in the Tender Notice No: NIT No. : IACS/CSS/BBD/2017-18/05, remain unchanged

Registrar

02.05.2017