

ASSOCIATION FOR THE CULTIVATION OF SCIENCE
2A & 2B Raja S.C. Mullick Road, Jadavpur, Kolkata - 700032

Ref. No.-IACS/CSS/PK/20-21/42-1

DATE : 18.03.2021

CORRIGENDUM NOTICE No.- III

NIT No. : IACS/CSS/PK/20-21/42

DATE: 03.03.2021

Following typo corrections in the corrigendum II must be taken care for the tender submission:

1. 2. Scanner

Existing: The Z scan range should be at least **10 μ m** or better, **and** optional up to 25 μ m. The Z scanner must have topography noise of less than 0.06 nm or better (bandwidth: 1 kHz)

Modified: The Z scan range should be at least **7 μ m** or better, **or** optional up to 25 μ m. The Z scanner must have topography noise of less than 0.06 nm or better (bandwidth: 1 kHz)

2. 5. Sample Size

Existing: Accommodate large samples at least **70 x 70 x 10 mm** (xyz) or better, must come with autotomized fine approach of z-stages (No third party motorized stage should be provided).

Modified: Accommodate large samples at least **40 x 40 x 10 mm** (xyz) or better, must come with autotomized fine approach of z-stages (No third party motorized stage should be provided).

-Praveen Kumar

School of Materials Science