

INDIAN ASSOCIATION FOR THE CULTIVATION OF SCIENCE

2A & 2B , Raja S.C. Mullick Road, Jadavpur, Kolkata-700032

NIT No. : IACS/CAM/NRJ/2017-18/72

Date: 20/11/ 2017

Sealed tender in two bids system (technical bid and price bid) is invited from bonafide, resourceful and eligible manufacturer/exclusive distributor/vendors for **“Gel Documentation System”**

Part-I (Technical Bid) of the tender should contain technical specifications in detail as well as commercial terms and conditions. Part-II (Price Bid) should clearly indicate group-wise price, if needed, as mentioned in the technical bid. The Technical Bid and Price Bid are to be submitted in separately sealed envelopes, distinctly marked accordingly and both to be put inside another envelop, that should be sealed and superscribed with tender notice no. and due date. The bidders may submit bids duly signed in their own letterheads.

Completed tender bids should reach the office of the **Registrar, Indian Association for the Cultivation of Science (IACS), 2A & 2B Raja S. C. Mullick Road, Jadavpur, Kolkata-700032** on or before the scheduled date and time specified below:

Tender Notice No.	NIT No. IACS/ CAM/NRJ// 2017-18 /72 Date: 20-11-17
Last date and time of submitting tender	December 15, 2017 at 1.00 p.m
Pre-bid meeting(date, time & place) to discuss technical specification	December 4, 2017 at 1.30 p.m, Centre for Advanced Materials Office
Date and time of opening tender	December 15, 2017 at 4.30 p.m
Place of opening tender	Centre for Advanced Materials Office
Date and time of opening of Price Bid	The Price Bids of the bidders qualifying the technical bid will only be opened, the date of which will be intimated to the short-listed bidders at their email addresses. The rest of the bids will be rejected.
Contact	Email: camnrj@iacs.res.in Tel.+91 33 2473 4971 [Ext 1195]

The technical bids will be opened first to judge/evaluate the technical specifications of the said instrument and thereafter the price bids of only technically qualified bidders will be opened.

Technical Bid Evaluation: The Technical Bids will be evaluated in the presence of the representatives of intending bidders who will be able to clarify technical aspects of their bids, if any, required by the Technical Evaluation Team.

Opening of price-bid: The Price Bids of the bidders qualifying the technical bid will only be opened, the date of which will be intimated to the short-listed bidders at their email addresses. The rest of the bids will be rejected.

Please note that IACS will not provide any accommodation or reimburse any expenses to any of the bidders for attending opening of technical bid.

Quotations received incomplete or beyond the stipulated time will be summarily rejected.

Bidders should submit their past experience for supplying and successful installation of similar units to other research Institutes/Universities/other organizations in India and abroad. Please provide documentary proofs of such successful installation and supportive documents that the instruments are running successfully.

1. TECHNICAL BID

The technical bid should contain technical specifications and should be kept in a separate sealed envelope duly superscribed as 'TECHNICAL BID' on the outer cover of the envelope as already detailed above. **It should be clearly mentioned on the envelope as "Technical Specification for Gel Documentation System".**

Technical Specification for "Gel Documentation System"

- Instrument should detect automatically (focusing and acquisition time) chemiluminescence, Fluorescence, & UV transilluminator.
- System with true 16 bit CCD camera; pixel density of 65,536 gray levels.
- Image resolution > 4 megapixel.
- Dynamic range should be at least 4 orders of magnitude.
- The camera should have peltier based cooling of minimum -30°C Absolute or -50°C from Room Temperature.
- Minimum sample size accommodated should be 28x36 cm.
- The system should be capable of measuring all possible dyes.
- Software should be PC and Mac compatible.
- System should be able to run in 15-35degC ambient temperature.
- System should be able to run between 100-240V and 50/60Hz power outlet.
- System should be able to determine molecular weight and display information with the band.
- Software can perform molecular weight determination, image integration etc.

Mini Vertical gel Specification- for 8 X 7 cm gels:

- High throughput- Capable of running up to 4 mini gel (8 X 7 Cm) simultaneously.
- Flexible- Capable of running hand cast as well as precast gel.
- Running and casting module should be different.
- Interchangeable module should be capable of using blotting module to do western blotting.
- Leak proof, tape free and easy assembly.
- Patented Flap wing for leak proof assembly.
- Permanently bonded spacer plates for leak proof, without agarose sealing & taping casting of gels.
- Casting frame with simple cam closure mechanism that gives precision alignment on any flat surface.
- Side by side casting stands that allow access to both gels simultaneously.
- Patented colored sample loading guides to prevent the skipping or repeated loading lanes.
- Modular design can be used to do western blotting by using the blotting module only.

- Should be able to run gels in 15-20 minutes.
- Should come with buffer dam.
- It should be supplied with 10% Stain free Fast Acrylamide Starter Kit.

Blotting Module:

- It should be compatible with small SDS page Unit.
- It should be capable of doing the western blotting of the mini gels.
- It can run 2 blots simultaneously.
- Color coded cassettes for proper orientation of gels and membranes.
- System is able to transfer two mini gels.
- Ice free cooling unit for proper temperature control for western blotting.
- Modular design can be used to do electrophoresis by using the electrode assembly and caster only.

Power Supply Basic:

- Programmable power supply should be capable to operate four electrophoresis units simultaneously for four identical runs with graphic LED display.
- The output range should be 10-300 V, 0.4-400 mA, 1-75 W.
- Constant voltage, current or Power with Automatic crossover.
- Memory storage: 9 programs, 9 steps, Timer Control: 99 h, 59 min.
- Automatic Power up after Power failure, Safety features: No-load detection; sudden load change detection.

Bio Dot SF Apparatus:

- Resistance to 100% ethanol, strong acid, and NaOH.
- Autoclavability.
- Patented sealing gasket to eliminate lateral leakage.
- Easy sample application with microplate-based spacing.
- 3-way flow valve for vacuum adjustment.

Important notes:

- i) Bidder providing misleading or wrong information will be disqualified
- ii) All technical claims of the Bidder should be supported by product catalogue, public website of the manufacture.
- iii) The Indian Agents desiring to quote on behalf of their foreign principals must have valid registration with the Central Purchase Organization (eg. DGS&D) as provided in Rule 143 of general Financial rules, 2005.

Warranty: Minimum 1 year while 3 years extendable warranty is preferred.

Servicing facility within India is a must while service engineer stationed at Kolkata will be preferred.

A list of other places where the instrument has been installed should also be provided.

A compliance table (see below) must be prepared and submitted along with the technical bid.

A compliance table (see below) must be prepared and submitted along with the technical bid.

Sr. No	Tender specification	Your offered instrument specification	Extent of compliance

2. PRICE BID

The financial bid indicating (item-wise) price for the item(s) mentioned in the technical bid should be kept in a separate sealed envelope duly super scribed as 'PRICE BID' on the outer cover of the envelop as already detailed above. Price bids of only technically qualified bidders will be opened and they will be intimated the date and time of the opening of price bid at their e-mail ids. Rest of the bidders will stand rejected.

PRICE: Price to be quoted on CIF Kolkata and also FOB basis.

3. BID SECURITY:

a. An Account payee Demand Draft/Pay Order of Nationalized Bank for Rs. **60,000/- (Rupees sixty thousand only)** drawn in favour of "Indian Association for the Cultivation of Science (State Bank of India, Jadavpur University Branch, A/C No. 11079699211, IFSC: SBIN000093, MICR Code: 700002048)" is to be furnished by the bidders alongwith Technical Bid except those who are registered with the Central Purchase Organizations, National Small Industries Corporation or the concerned Ministry or Department, as Bid Security money or Earnest Money Deposit (EMD). Alternatively, the Bidder shall have the option to furnish the EMD in the format of Bank Guarantee (BG).

b. The Demand Draft for the Bid-Security should have at least about 90 (ninety) days validity period of opening of the bids.

c. In case of non-award of the work the Bid Security money would be returned to the unsuccessful Bidders.

4. PERFORMANCE SECURITY:

An Account Payee Demand Draft on any nationalized bank of India of 10% of the order value in the name of "Indian Association for the Cultivation of Science" is to be furnished by the successful bidder as Performance security. Performance security money should remain valid for a period of 60 days beyond the date of completion of all contractual obligations of the supplier including warranty obligations. Bid security money or EMD will be refunded to successful bidder on receipt of the Performance security money.

5. TERMS OF PAYMENT :

Payment will be made through irrevocable Letter of Credit in two installments. 90% of the money will be released on submission of shipping of documents. Remaining 10% will be released after successful installation of the instrument.

6. GENERAL INSTRUCTIONS :

1. Validity of tender: Tender submitted should remain valid for at least six months from the date of opening the tender. Validity beyond six months from the date of opening of the tender shall be lapsed by mutual consent.
2. The tender should accompany a compliance chart.
3. Incomplete and conditional tenders as well as tenders received after the due date will be summarily rejected without assigning any reasons thereof.
4. At any time prior to the bid due date, IACS may, for any reason, whether at its own initiative or in response to a clarification requested by a prospective bidder during pre-bid meeting, modify the bidding documents. The amendment(s) will be notified on the institute website. Prospective bidders are advised to occasionally to visit the website (www.iacs.res.in/tender) for any amendment.
5. Installation/Demonstration/Application training at site: Installation & user training at IACS, free of cost by the supplier.
6. Service facility: In India, preferably Kolkata, supplier should mention their details of service setup and man powers who are responsible for after sales support. Response time should be within 24 hrs.
7. The model number, make and a printed literature of the product should be submitted positively.
8. Proposed delivery schedule should be mentioned clearly.
9. Manufacturers / exclusive distributors / vendors should have history of supplying this type of instruments to this or other scientific organizations. Availability of a list in this regard would be preferred.
10. Authorized dealership certificate should be provided in case of principal manufacturing company is not quoting directly.
11. Guarantee certificate, users manuals etc. are to be handed over to the user after successful commissioning of the system.
12. In the event of date being declared a closed holiday for purchaser's office, the due date for submission of bids and opening of the technical bids will be the following working day at the appointed time.
13. In case of any dispute, the decision of IACS authority shall be final and bidding on the bidders.
14. For any clarification regarding technical specifications, information etc., please send your queries to camnrj@iacs.res.in.
15. The authority of IACS reserves the right to reject any or all of the tenders received without assigning any reason thereof.

Acting Registrar