

INDIAN ASSOCIATION FOR THE CULTIVATION OF SCIENCE  
2A & 2B Raja S.C. Mullick Road, Jadavpur, Kolkata – 700032

Ref. No.- SCS/TC/19-20/102-1

DATE: 12-02-2020

**CORRIGENDUM NOTICE No.- I**

**NIT No. IACS/SCS/TC/2019-20/102**

**DATE: 29-01-2020**

Based on the discussion in the pre-bid meeting held on February 07, 2020 at Lakeview Academic Building Meeting Room, the technical committee modified few specifications of 'Robust UG-PG Teaching Lab FTIR Spectrometer' as specified below :-

**Technical Specification for Research Grade FTIR Spectrometer: (1) Should be robust for use under the ambient atmospheric conditions of a teaching lab (2) Spectral range: mid-IR (400-6000  $\text{cm}^{-1}$ ) (3) Spectral resolution  $<1\text{cm}^{-1}$  (4) Accessory: For Gaseous sample analysis.**

**Acting Registrar**