

Corrigendum – I

Tender NIT No. : IACS/SMS/UM/21-22/62 dated 22.07.2021

Following corrections to the tender document were agreed upon in the pre-bid meeting.

Under section “Technical specification for Inert Atmosphere Glove box”

Point no. 2. To be changed to: Inner Box Dimensions in the range: 900-950 mm x 1150-1250 mm x 725-800 mm [H x L x D]. Inside and outside surfaces should be completely corrosion resistant.

Under section “Optional Items”

Point “Solvent adsorption unit Activated carbon, at least 5kgs. Should be with inline and bypass mode.” to be changed to “Activated carbon, at least 5kgs for solvent adsorption unit”.

Registrar