

INDIAN ASSOCIATION FOR THE CULTIVATION OF SCIENCE
2A & 2B, Raja S.C.Mullick Road, Jadavpur, Kolkata-700032

Corrigendum-1

IACS/CSS/SV/121 Date: 25-01-2022,

Please find the following changes in the tender notice of **Compact fluorescence spectrometer for steady state spectral measurements**.

Existing Specification	Change / Insertion
Detailed specifications: Solid state sample holder with variable angle and preferably linear adjustment for focusing the excitation on the solid-state sample	Detailed specifications: Additional preference will be given for the spectrometer with linear and rotational stage for solid state sample holder.
Photomultiplier tube with peltier cooled housing	Detailed specifications: PMT with either peltier or air-cooled systems will be considered, if a signal to noise ratio of 10,000:1 FSD of water Raman peak can be delivered.

All other terms and conditions/specifications remain unchanged.

Acting Registrar